

## How to create consistent comparative response curves of Cochlear Implant microphones



1. Cochlear Nucleus Freedom BTE – DCTEST2 Lead only. 2. Cochlear Nucleus 5 CP810 BTE – DCTEST3 Lead only.
3. Cochlear Nucleus 6 CP910 BTE – DCTEST3 Lead only. 4. Advanced Bionics Harmony – DCTEST4 lead and AB Harmony Listening Check. CI-5821.
5. Advanced Bionics Neptune – DCTEST4 lead and Neptune Connect AB: CI-5241-WHT (White)-BLK (Black).
6. Advanced Bionics Naida Q70 BTE – DCTEST4 lead and Naida Listening Check. CI-5823.
7. Med-EL Opus2 BTE - DCTEST4 lead and Microphone Test Device 02883, Mini Battery Pack (28cm lead) 08264, MBPMTD Connecting Cable 08082.
8. Med-EL Rondo – DCTEST4 lead and Microphone Test Device 02883, Mini Battery Pack with Rondo Connection cable, MBPMTD C.Cable 08082.
9. Med-EL Sonnet BTE – DCTEST4 lead and Microphone Test Device 02883, Mini Battery Pack lead (currently unavailable) MBPMTD C.Cable 08082.